PATENT COOPERATION TREATY

From the INTERNATIONAL BUREAU PCT Commissioner **NOTIFICATION OF ELECTION US Department of Commerce** United States Patent and Trademark (PCT Rule 61.2) Office, PCT 2011 South Clark Place Room CP2/5C24 Arlington, VA 22202 **ETATS-UNIS D'AMERIQUE** Date of mailing (day/month/year) in its capacity as elected Office 29 May 2001 (29.05.01) International application No. Applicant's or agent's file reference PCT/AU00/01038 40455407 International filing date (day/month/year) Priority date (day/month/year) 01 September 2000 (01.09.00) 02 September 1999 (02.09.99) **Applicant** GRAYDEN, David, Bruce et al 1. The designated Office is hereby notified of its election made: in the demand filed with the International Preliminary Examining Authority on: 15 March 2001 (15.03.01) in a notice effecting later election filed with the International Bureau on: 2. The election was not made before the expiration of 19 months from the priority date or, where Rule 32 applies, within the time limit under Rule 32.2(b).

The International Bureau of WIPO 34, chemin des Colombettes 1211 Geneva 20, Switzerland

Authorized officer

Claudio Borton

Telephone No.: (41-22) 338.83.38

Facsimile No.: (41-22) 740.14.35

(19) World Intellectual Property Organization International Bureau



(43) International Publication Date 15 March 2001 (15.03.2001)

PCT

(10) International Publication Number WO 01/19135 A1

- (51) International Patent Classification7: A61F 2/18
- H04R 25/00.
- (21) International Application Number: PCT/AU00/01038
- (22) International Filing Date:

1 September 2000 (01.09.2000)

(25) Filing Language:

English

(26) Publication Language:

English

(30) Priority Data:

PQ 2612

2 September 1999 (02.09.1999) AU

- (71) Applicant (for all designated States except US): THE BIONIC EAR INSTITUTE [AU/AU]; 384-388 Albert Street, East Melbourne, VIC 3002 (AU).
- (72) Inventors; and
- (75) Inventors/Applicants (for US only): GRAYDEN, David, Bruce [AU/AU]; 16 Dresden Avenue, Heathmont, VIC 3135 (AU). CLARK, Graeme, Milbourne [AU/AU]; 13 Banoon Road, Eltham, VIC 3095 (AU).

- (74) Agent: FREEHILLS CARTER SMITH BEADLE; Level 47, 101 Collins Street, Melbourne, VIC 3000 (AU).
- (81) Designated States (national): AE, AG, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, BZ, CA, CH, CN, CR, CU, CZ, DE, DK, DM, DZ, EE, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MA, MD, MG, MK, MN, MW, MX, MZ, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, TZ, UA, UG, US, UZ, VN, YU, ZA, ZW.
- (84) Designated States (regional): ARIPO patent (GH, GM, KE, LS, MW, MZ, SD, SL, SZ, TZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TI, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG).

Published:

With international search report.

For two-letter codes and other abbreviations, refer to the "Guidance Notes on Codes and Abbreviations" appearing at the beginning of each regular issue of the PCT Gazette.

(54) Title: IMPROVED SOUND PROCESSOR FOR COCHLEAR IMPLANTS

(57) Abstract: An improved sound processor for a cochlear implant having electrodes for stimulating the auditory nerve, including means for receiving sounds, means for processing the sounds and converting them to electrical stimulation signals for application to the electrodes of the cochlear implant for stimulation of the auditory nerve, said sound processing means including means for generating electrical signals to be applied to the basal electrodes having different predetermined rates of stimulation and the implant having basal electrodes and apical electrodes and the means for generating electrical signals to be applied to the apical electrodes have a different rate of stimulation, the electrical signals to be applied to the basal electrodes having a higher rate of stimulation than the electrical signals to be applied to the apical electrodes.

15

20

10/069846 JC13 Rec'd PCT/PTO 28 FEB 2002 PCT/AU00/01038

-1-

IMPROVED SOUND PROCESSOR FOR COCHLEAR IMPLANTS

Field of the Invention

This invention relates to improvements in sound processors for cochlear implants, and more particularly to a Differential Rate Sound Processor (DRSP)

Background of the Invention

The multi-channel cochlear implant was first implanted in 1978. Early signal processing designs extracted the second formant (F2) and pitch (F0) to control electrode stimulation. The frequency of F2 controlled the location of electrode stimulation, and F0 controlled the rate of stimulation. This was later improved by also extracting the first formant (F1) and adding a second stimulated electrode for each pitch period. The MULTI-PEAK (MPEAK) stimulation strategy added stimulation of a number of fixed electrodes to better represent high-frequency information. The next stages of development were the SMSP and SPEAK strategies. These were a departure from the others at they used a fixed stimulation rate and stimulated electrodes that corresponded to maxima in the sound spectra. Another fixed-rate strategy, CIS, was developed overseas. This strategy stimulated all of a small number of electrodes to represent the sound spectra. All of the above processing strategies involve fixed-rate sound processing.

The named inventors have determined that some speech features are better perceived using low-rates of simulation, while some are better perceived using high rates of stimulation. Higher rates of stimulation present more information about phonetic manner of articulation, but spectral information tends to be smeared at such higher rates.

Summary of the Invention and Object

It is an object of the present invention to provide an improved sound processor for use with cochlear implants in which the problems associated with fixed rate stimulation are ameliorated.

The invention provides in one form an improved sound processor for a cochlear implant having electrodes for stimulating the auditory nerve, including means for receiving sounds, means for processing the sounds and converting them to electrical stimulation signals for application to the electrodes of the cochlear implant for stimulation of the auditory nerve, said sound processing means including means for generating electrical signals to be applied to the electrodes having different predetermined rates of stimulation.

In this first form of the invention, the cochlear implant preferably has basal electrodes and apical electrodes and the means for generating electrical signals to be applied to the apical electrodes have a different rate of stimulation, the electrical signals to be applied to the basal electrodes having a higher rate of stimulation than the electrical signals to be applied to the apical electrodes.

By causing stimulation of the basal electrodes at a higher rate of stimulation than the apical electrodes, the manner of articulation features of speech will be more optimally presented to the cochlear implant user, leading to improved speech understanding performance. High rates of stimulation at the basal electrodes will present good information about temporal events and frication. The low rates of stimulation of the apical electrodes will present good spectral information in this regard, where most place of articulation features reside.

In a preferred embodiment, the more apical electrodes will be chosen as those that contain the voice bar and lower formants of speech. In this frequency region, spectral detail is important and the apical electrodes will be stimulating using a stimulation rate of between about 250 cycles per second and about 800 cycles per second, depending on the user. By adopting stimulation rates falling within the above range, better information about place of articulation of speech, which is largely represented by the formant structure, is obtained by the user.

The more basal electrodes represent higher frequency components of the incoming sound, and higher rates of stimulation of these electrodes will be used

to better represent noise and more precisely present information about temporal events such as rapid changes in amplitude. The latter is important for perception of manner of articulation and voicing. These electrodes will be stimulated at a higher rate than the apical electrodes, with stimulation rates at or above about 800 cycles per second, and preferably up to about 1600 cycles per second, being selected depending on the user.

In the case of an implant having 20 electrodes available for stimulation, the apical electrodes are electrodes 0 to 12, and the basal electrodes are electrodes 13 to 19. The apical electrodes represent sound frequencies from 0 to about 2700Hz, while the basal electrodes represent frequencies from about 2700Hz to about 7900Hz. The stated apical electrode frequencies are sufficient to contain the first three formants of most speech.

In a particularly preferred form of the invention, the apical electrodes are stimulated at about 250 cycles per second while the basal electrodes are stimulated at about 1500 cycles per second. To ensure that stimulation levels are suitable for these different rates, the threshold (T) levels and comfort (C) levels of the patient are carefully set. The electrodes to be stimulated are chosen by selecting the eight largest spectral energies within filterbanks derived from the Fast Fourier Transform (FFT) or the Discrete Wavelet Transform (DWT) which is performed by the processor.

In another form, the invention provides an improved sound processor for a cochlear implant having electrodes for stimulating the auditory nerve, including means for receiving sounds, means for processing the sounds and converting to electrical stimulation signals for application to the electrodes of the cochlear implant whereby the auditory nerve is electrically stimulated, said sound processing means having means for varying the rate of stimulation of the electrical stimulation signals depending on the parameters of the sound received by the sound receiving means.

10

15

By varying the rate of stimulation of the cochlear implant electrodes depending on the incoming speech signal, key speech features will be more optimally presented to the cochlear implant user thereby leading to improved speech understanding performance.

In a preferred form of this aspect of the invention, the sound processing means will be programmed to continually adjust the rate of stimulation of the electrical stimulation signals depending on the parameters of the incoming speech signal. To this end, the incoming speech signal will be processed to detect events that are better represented using a higher rate of stimulation. Such events include plosive onset bursts, frication and other rapid spectral changes. The rate of stimulation across all electrodes will be increased for the average duration of these events. The standard rate will be between 250 cycles/s and 800 cycles/s depending on the user. The higher rate will be above about 800 cycles/s, and preferably up to about 1600 cycles/s, also depending on the user.

In order that the invention may be more readily understood, one presently preferred embodiment of the invention will now be described.

Description of Preferred Embodiment

The invention is preferably designed for use with the CI-24M Cochlear Implant as manufactured by Cochlear Ltd, and as described in US Patent No. 4532930, the contents of which are incorporated herein by cross-reference, and later patents by Cochlear Ltd to be found in the patent literature.

Although the CI-24M Implant will be used in most cases, the invention could be applied to any implant that uses pulsatile stimulation. The stimulation strategy is based on the Spectral Maxima Sound Processor (SMSP), which is described in United States Patent 5597380 and Australian Patent 657959, the contents of which are incorporated herein by cross reference although other strategies may be used with similar results. For example, the SPEAK strategy as discussed in US Patent No. 5,597,380, the contents of which are incorporated by cross reference.

The electrode selection strategy from the SMSP is varied to ensure that electrodes are stimulated at the desired predetermined frequencies for each cycle of stimulation. The preferred signal processing device will be the SPEAR processor, which is currently under development at The Bionic Ear Institute, and which is described in the following paper:

Zakis, J.A. and McDermott, H.J. (1999). "A new digital sound processor for hearing research," Proceedings of the Inaugural Conference of the Victorian Conference of the Victorian Chapter of the IEEE Engineering in Medicine and Biology Society, February 22-23, pp. 54-57. The contents of this paper are similarly incorporated herein by cross reference.

The processor is a generic processor based on the Motorola DSP56300 family, such as the DSP56302, or the DSP56309, although any digital signal processor, including those produced by Cochlear Ltd and their competitors, could be used to run the differential rate sound processor program of the present invention, provided they have adequate processing speed.

In the implementation of the first form of the invention, the differential rate stimulation processor software embodying the invention is downloaded to the SPEAR processor and stored on EPROM. Patient map details, including frequency bands, threshold (T) levels and comfort (C) levels, are also stored on the device. Monopolar stimulation mode is used to reduce current levels and for longer battery life.

For the case where 20 electrodes are available for stimulation, the apical electrodes are electrodes 0 to 12, and the basal electrodes are electrodes 13 to 19. The apical electrodes then represent frequencies from 0 to 2700Hz; the basal electrodes represent frequencies from 2700Hz to 7900Hz. The stated apical electrode frequencies are sufficient to contain the first three formants of most speakers' speech.

The apical electrodes are stimulated at about 250 cycles/s and the basal electrodes at about 1500 cycles/s. The patient's T and C levels are carefully set

10

15

20

25

to ensure that stimulation levels are suitable for the two different rates and adjustments made if necessary. The electrodes to be stimulated are chosen by selecting the eight largest spectral energies within filterbanks derived from the Fast Fourier Transform (FFT) or the Discrete Wavelet Transform (DWT).

The values quoted above are examples. Patient-to-patient variability is large and some need higher stimulation rates on the apical electrodes and/or lower stimulation rates on the basal electrodes. These are determined for each individual by evaluating a number of rate combinations in every day usage. Also, some patients do not have as many electrodes available and so the choice of electrodes is altered to suit their situation. However, the spectral ranges of the apical and basal electrodes remain much the same.

By using the Differential Rate Sound Processor (DRSP) program of the invention, features of speech will be more optimally presented to the cochlear implant user leading to improved speech understanding performance.

In the implementation of the second aspect of the invention, the software necessary to provide a variable rate of stimulation depending on the incoming speech signal is downloaded to the SPEAR processor and stored on an EPROM. Patient map details, including frequency bands, threshold (T) levels and comfort (C) levels, are also stored on the device. Monopolar stimulation mode is used to reduce current levels and for longer battery life.

The standard rate of stimulation is about 250 cycles/s and the higher rate is about 1500 cycles/s. The patient's T and C levels are carefully set to ensure that stimulation levels are suitable for the two different rates. The electrodes to be stimulated are chosen by selecting the eight largest spectral energies within filterbanks derived from the Fast Fourier Transform (FFT) or the Discrete Wavelet Transform (DWT).

The changes in spectral energies and the amount of frequency energy are monitored over time. When there is a significantly large change between frames separated by the period of the lower stimulation rate then the higher stimulation

rate is used for 50 ms. This procedure locates plosive bursts and other rapid spectral changes. The higher stimulation rate is also used when the ratio of energy below about 300Hz to that above about 2000Hz is less than about 0.5. This locates phonemes with significant frication.

The values quoted above are examples. Patient-to-patient variability is large and some need a higher stimulation rate for the standard rate and/or a lower stimulation rate for the higher rate. These are determined for each individual by evaluating a number of rate combinations in every day usage. Thresholds for changes in energy and ratio of energies are also adjustable for each individual.

CLAIMS:

1. A sound processor for a cochlear implant having electrodes for stimulating the auditory nerve, including means for receiving sounds, means for processing the sounds and converting them to electrical stimulation signals for application to the electrodes of the cochlear implant for stimulation of the auditory nerve, said sound processing means including means for generating electrical signals to be applied to the basal electrodes having different predetermined rates of stimulation.

10

- 2. The sound processor of claim 1, wherein the cochlear implant has one form of the invention, the cochlear implant has basal electrodes and apical electrodes and the means for generating electrical signals to be applied to the apical electrodes have a different rate of stimulation, the electrical signals to be applied to the basal electrodes having a higher rate of stimulation than the electrical signals to be applied to the apical electrodes.
- 3. The sound processor of claim 2, wherein the more apical electrodes are selected for stimulator signals that represents the voice bar and lower formants of the sounds.
 - 4. The sound processor of claim 3, where the more apical electrodes apply stimulation signals having a stimulation rate of between about 250 cycles per second and about 800 cycles per second depending on the user, to provide precise spectral and place of articulation information.
 - 5. The sound processor of any one of claims 2 to 4, wherein the more basal electrodes apply stimulation signals having a stimulation rate of at or above

15

about 800 cycles per second depending on the user, to provide precise information about temporal events and frication.

- 6. The sound processor of any one of claims 2 to 5, including a twenty (20) electrode implant, the apical electrodes are electrodes 0 to 12 and the basal electrodes are electrodes 13 to 19, the apical electrodes representing sound frequencies from 0 to about 2700Hz, while the basal electrodes represent frequencies from about 2700Hz to about 7900Hz, the apical electrode frequencies, are sufficient to contain the first three formants of most speakers speech.
- 7. The sound processor of any one of claims 2 to 6, wherein the apical electrodes are stimulated at about 250 cycles per second while the basal electrodes are stimulated at about 1500 cycles per second.
- 8. The sound processor of claim 7 wherein the threshold (T) levels and comfort (C) levels of the patient are carefully set, the electrodes to be stimulated are chosen by selecting the eight largest spectral energies within filterbanks derived from the Fast Fourier Transform (FFT) or the Discrete Wavelet Transform (DWT) which is performed by the processor.
- 9. A sound processor for a cochlear implant having electrodes for stimulating the auditory nerve, including means for receiving sounds, means for processing the sounds and converting to electrical stimulation signals for application to the electrodes of the cochlear implant whereby the auditory nerve is electrically stimulated, said sound processing means having means for varying the overall rate of stimulation of the electrical stimulation signals depending on the parameters of the sound received by the sound receiving means.

- 10. The sound processor of claim 9, wherein the sound processing means is programmed to continually adjust the rate of stimulation of the electrical stimulation signals depending on the parameters of the incoming speech signal, to this end, the incoming speech signal will be processed to detect events that are better represented using a higher rate of stimulation.
- 11. The sound processor of claim 9 or 10, as defined in any one of claims 2 to 8.
- 10 12. The sound processor of any preceding claim, wherein the implant is one which uses pulsatile stimulation.
 - 13. The sound processor of any preceding claim wherein a SPEAR processing device is programmed using the Differential Rate Sound Processor (DRSP) program defined in any preceding claim to optimally present the features of speech to the implant.

INTERNATIONAL SEARCH REPORT

International application No.

PCT/AU 00/01038

A. (CLASSIFICATION OF SUBJECT MATTER				
Int Cl ⁷ :	H04R 25/00; A61F 2/18				
According to International Patent Classification (IPC) or to both national classification and IPC					
		CRASSILISATE IN THE STATE OF TH			
	FIELDS SEARCHED				
IPC: H04R,					
Documentation AU; IPC as a	searched other than minimum documentation to the externove	at that such documents are included in the	e fields searched		
DERWENT: JAPIO :	base consulted during the international search (name of de cochlear, biomic ear, inside ear, sound, voice, p same as above same as above	ata base and, where practicable, search te process, electrode, stimulation, ex	ams used) citation		
C.	DOCUMENTS CONSIDERED TO BE RELEVANT				
Category*	Citation of document, with indication, where appr	opriate, of the relevant passages	Relevant to claim No.		
	Journal of the Acoustical Society of America, Volume 98, No. 2, pt. 1 issued August 1995, Sipke Pijl and Dietrich W.F. Schwarz, "Melody recognition and musical interval perception by deaf subjects stimulated with electric pulse trains through single cochlear implant electrodes" Pages 886 - 895				
Y	•	71397 A(SELIGMAN et al.), 21 December 1993 e document 1/03913 A1 (COCHLEAR PTY LIMITED), 21 March 1991			
Y	Whole document WO 91/03913 A1 (COCHLEAR PTY LIMIT Whole document				
. X Further documents are listed in the Continuation of Box C					
Special categories of cited documents: "A" Document defining the general state of the art which is not considered to be of particular relevance "E" carlier application or patent but published on or after the international filing date "L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified) "O" of other means "P" document published prior to the international filing date "L" document of particular relevance, the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone document referring to an oral disclosure, use, exhibition or other means "P" document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such document, such combination being obvious to a person skilled in the art document member of the same patent family					
Date of the actual completion of the international search Date of mailing of the international search 29 SEP 2000					
22 Sopulation 2000					
AUSTRALIA PO BOX 200 WODEN AG E-mail addi	uling address of the ISA/AU AN PATENT OFFICE CT 2606 AUSTRALIA ress: pct@ipaustralia_gov.au a.: (02) 6285 3929	MANISH RAJ Telephone No.: (02) 6283 2175			

INTERNATIONAL SEARCH REPORT

L .national application No. PCT/AU 00/01038

PC1/AU 00/01038						
C (Continuation). DOCUMENTS CONSIDERED TO BE RELEVANT						
Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.				
	Journal of Acoustical Society of America, Volume 98, No. 4, October 1995, David A. Nelson, Dianne J. Van Tasell, Anna C. Schroder, Sigfrid Soli, Samuel Lerine "Electrode ranking of place pitch and speech recognition in					
4	electrical hearing", pages 1987-1999. Abstract	1				
A	US 5922016 A (WAGNER), 13 July 1999	1-13				
A	Whole document	1-15				
A	US 5749912 A (ZHANG et al), 12 May 1998 Whole document	1-13				
		·				
	·					

INTERNATIONAL SEARCH REPORT

Information on patent family members

International application No. PCT/AU 00/01038

This Annex lists the known "A" publication level patent family members relating to the patent documents cited in the above-mentioned international search report. The Australian Patent Office is in no way liable for these particulars which are merely given for the purpose of information.

Patent Do	cument Cited in Search Report			Pateru	Family Member		
US	5 271 397	ΑÜ	63392/90	CA	2 024 845	EP	450 004
		wo	91/03 913	US	5 095 904		
wo	91/03 913	AU	63392/90	CA	2 024 845	EP	450 004
		US	5 271 397	US	5 095 904		
US	5 749 912	AU	38899/95	US	5 549 658	wo	9 612 456

END OF ANNEX

INTERNATIONAL PRELIMINARY EXAMINATION REPORT

(PCT Article 36 and Rule 70)

Applicant's or agent's file reference FOR FURTHER See Notification of Transmittal of International Prelimin ACTION Examination Report (Form PCT/IPEA/416).		ransmittal of International Preliminary (Form PCT/IPEA/416).			
International Application No.	International Filing Da	ate (day/month/year)	Priority Date (day/month/year)		
PCT/AU00/01038	1 September 2000		2 September 1999		
International Patent Classification (IPC)	International Patent Classification (IPC) or national classification and IPC				
Int. Cl. 7 H04R 25/00; A61F 2/18					
Applicant					
THE BIONIC EAR INSTITU	TE et al				
This international proliminary and is transmitted to the appli	1. This international proliminary examination report has been prepared by this International Preliminary Examining Authority and is transmitted to the applicant according to Article 36.				
2. This REPORT consists of a to	tal of 3 sheets, inclu-	ding this cover sheet.			
This report is also accor	manied by ANNEXES.	i.e., sheets of the descri	iption, claims and/or drawings which have		
been amended and are the	This report is also accompanied by ANNEXES, i.e., sheets of the description, claims and/or drawings which have been amended and are the basis for this report and/or sheets containing rectifications made before this Authority (see Rule 70.16 and Section 607 of the Administrative Instructions under the PCT).				
Rule /0.16 and Section	60 / 61 the Administrativ	ve instructions under the	ercij.		
These annexes consist of a tot	These annexes consist of a total of sheet(s).				
3. This report contains indications relat	3. This report contains indications relating to the following items:				
I X Basis of the repo	I X Basis of the report				
II Priority					
III Non-establishme	Non-establishment of opinion with regard to novelty, inventive step and industrial applicability				
V Lack of unity of	invention				
V X Reasoned statem citations and exp	ement under Article 35(2) with regard to novelty, inventive step or industrial applicability; explanations supporting such statement				
VI Certain documen	its cited				
VII Certain defects in	n the international application				
VIII Certain observations on the international application					
Date 5 - Julian of the report					
Date of submission of the demand		Date of completion of the report			
15 March 2001		3 April 2001			
Name and mailing address of the IPEA/AU		Authorized Officer			
AUSTRALIAN PATENT OFFICE PO BOX 200, WODEN ACT 2606, AUST	TRALIA				
E-mail address: pct@ipaustralia.gov.au Facsimile No. (02) 6285 3929		MANISH RAJ			
1 acsume 10. (02) 0203 3723		Telephone No. (02) 6283 2175			

PCT/AU00/01038

I.	Basis of the report
1.	With regard to the elements of the international application:*
	X the international application as originally filed.
	the description, pages, as originally filed,
į	pages , filed with the demand,
	pages, received on with the letter of
	the claims, pages, as originally filed,
	pages, as amended (together with any statement) under Article 19,
	pages, filed with the demand,
	pages, received on with the letter of
	the drawings, pages, as originally filed,
	pages, filed with the demand,
	pages, received on with the letter of the sequence listing part of the description:
	pages, as originally filed
	pages, filed with the demand pages, received on with the letter of
2.	With regard to the language, all the elements marked above were available or furnished to this Authority in the language in which the international application was filed, unless otherwise indicated under this item.
	These elements were available or furnished to this Authority in the following language which is:
	the language of a translation furnished for the purposes of international search (under Rule 23.1(b)).
	the language of publication of the international application (under Rule 48.3(b)).
	the language of the translation furnished for the purposes of international preliminary examination (under Rules 55.2 and/or 55.3).
3.	With regard to any nucleotide and/or amino acid sequence disclosed in the international application, was on the basis of the sequence listing:
	contained in the international application in written form.
	filed together with the international application in computer readable form.
	furnished subsequently to this Authority in written form.
	furnished subsequently to this Authority in computer readable form.
	The statement that the subsequently furnished written sequence listing does not go beyond the disclosure in the international application as filed has been furnished.
	The statement that the information recorded in computer readable form is identical to the written sequence listing has been furnished
4.	The amendments have resulted in the cancellation of:
	the description, pages
	the claims, Nos.
	the drawings, sheets/fig.
5.	This report has been established as if (some of) the amendments had not been made, since they have been considered to go beyond the disclosure as filed, as indicated in the Supplemental Box (Rule 70.2(c)).**
*	Replacement sheets which have been furnished to the receiving Office in response to an invitation under Article 14 are referred to in this report as "originally filed" and are not annexed to this report since they do not contain amendments (Rules 70.16 and 70.17).
**	Any replacement sheet containing such amendments must be referred to under item I and annexed to this report

Claims

Claims

Claims 1 - 13

NO

YES

NO

PCT/AU00/01038

	and explanations supporting such statement			
1.	Statement			
	Novelty (N)	Claims 1 - 13	YES	
		Claims	NO .	
	Inventive step (IS)	Claims 1-13	YES	

2. Citations and explanations (Rule 70.7)

Industrial applicability (IA)

V.

- 1. Claims 1-13 are novel and involve inventive step because no individual citation or obvious combination of citations disclose a sound processor for a cochlear implant including means for receiving sounds, means for processing sounds and converting them to electrical stimulation signals for application to the electrodes of the cochlear implant and sound processing means including means for generating electrical signals to be applied to the basal electrodes having different predetermined rates of stimulation.
- 2. Claims 1-13 have industrial applicability because the invention claimed can be made or used in the industry.